

Human IL-8 ELISA Kit

Instructions for use

Catalogue numbers:

1x48 tests: 950.050.048
1x96 tests: 950.050.096
2x96 tests: 950.050.192

For research use only

Fast Track Your Research.....

Table of Contents

1.	Intended use.....	2
2.	Introduction.....	2
2.1.	Summary.....	2
2.2.	Principle of the method.....	3
3.	Reagents provided and reconstitution.....	3
4.	Materials required but not provided.....	4
5.	Storage Instructions	4
6.	Specimen collection, processing & storage	4
7.	Safety & precautions for use	5
8.	Assay Preparation	6
8.1.	Assay Design	6
8.2.	Preparation of Wash Buffer.....	6
8.3.	Preparation of Standard Diluent Buffer	6
8.4.	Preparation of Standard	7
8.5.	Preparation of Controls	7
8.6.	Preparation of Biotinylated anti-IL-8	8
8.7.	Preparation of Streptavidin-HRP	8
9.	Method	9
10.	Data Analysis.....	10
11.	Assay limitations.....	10
12.	Performance Characteristics.....	11
12.1.	Sensitivity	11
12.2.	Specificity	11
12.3.	Precision.....	11
12.4.	Dilution Parallelism.....	12
12.5.	Spike Recovery	12
12.6.	Stability	13
12.7.	Expected serum values	13
12.8.	Standard Calibration.....	13
13.	Bibliography.....	14
14.	Diaclone Human IL-8 ELISA references	14
15.	Assay Summary	15
16.	International Summaries	16
16.1.	French	16
16.2.	Spanish	18

Human IL-8 ELISA KIT

1. Intended use

The Diaclone IL-8 ELISA kit is a solid phase sandwich ELISA for the *in-vitro* qualitative and quantitative determination of IL-8 in supernatants, buffered solutions or serum and plasma samples. This assay will recognise both natural and recombinant human IL-8.

This kit has been configured for research only in the rest of the world. Not suitable for use in therapeutic procedures.

2. Introduction

2.1. Summary

Interleukin 8 (IL-8) or CXCL8, Monocyte-Derived Neutrophil Chemotactic Factor (MDNCF), Neutrophil Activating Factor (NAF) and NAD-P1 is a chemokine secreted by monocytes, macrophages and endothelial cells. IL-8 chemoattracts and activates neutrophils.(1, 2, 3)

The predominant form of IL-8 is a 8.4kDa protein containing 72 amino acid residues, which includes five additional N-Terminal amino-acids. IL-8 contains the four conserved cysteine residues present in CXC chemokines and also contains the "ELR" motif common to CXC chemokines that binds to CXCR1 and CXCR2.(3,4)

Data indicate that IL-8 may participate in the pathogenesis of rheumatoid arthritis (5) via the induction of neutrophil-mediated cartilage damage (6), and psoriasis(7). A causative involvement of IL-8 is found within a broad range of clinico-pathological conditions: adult respiratory distress syndrome, asthma, bacterial infections, bladder cancer, graft rejection and influenza infection, due to the now known biological properties of IL-8. This cytokine especially in combinations with other neutrophil activating agents, may prove helpful in the treatment of patients suffering from granulocytopenia, severe infections against which antibiotics are not effective, and immunodeficiency caused by HIV (8)

2.2. Principle of the method

A capture Antibody highly specific for IL-8 has been coated to the wells of the microtitre strip plate provided during manufacture. Binding of IL-8 samples and known standards to the capture antibodies and subsequent binding of the biotinylated anti- IL-8 secondary antibody to the analyte is completed during the same incubation period. Any excess unbound analyte and secondary antibody is removed. The HRP conjugate solution is then added to every well including the zero wells, following incubation excess conjugate is removed by careful washing. A chromogen substrate is added to the wells resulting in the progressive development of a blue coloured complex with the conjugate. The colour development is then stopped by the addition of acid turning the resultant final product yellow. The intensity of the produced coloured complex is directly proportional to the concentration of IL-8 present in the samples and standards. The absorbance of the colour complex is then measured and the generated OD values for each standard are plotted against expected concentration forming a standard curve. This standard curve can then be used to accurately determine the concentration of IL-8 in any sample tested.

3. Reagents provided and reconstitution

Reagents (Store @2-8°C)	Quantity 1x48 well kit Cat no. 950.050.048	Quantity 1x96 well kit Cat no. 950.050.096	Quantity 2x96 well kit Cat no. 950.050.192	Reconstitution
96 well microtitre strip plate	1/2	1	2	Ready to use (Pre-coated)
Plastic plate covers	2	2	4	n/a
Standard: 2000pg/ml	1	2	4	Reconstitute as directed on the vial (see reagent preparation, section 8)
Standard Diluent (Buffer)	1 (25ml)	1 (25ml)	1 (25ml)	10x Concentrate, dilute in distilled water (see reagent preparation, section 8)
Standard Diluent (Human serum)	1 (7ml)	1 (7ml)	2 (7ml)	Ready to use
Control	1	2	4	Reconstitute as directed on the vial (see reagent preparation, section 8)
Biotinylated anti-IL-8	1 (0.4ml)	1 (0.4ml)	2 (0.4ml)	Dilute in biotinylated antibody diluent (see reagent preparation, section 8)
Biotinylated Antibody diluent	1 (7ml)	1 (7ml)	1 (13ml)	Ready to use
Streptavidin-HRP	1 (5µl)	2 (5µl)	4 (5µl)	Add 0.5ml of HRP diluent prior to use (see reagent preparation, section 8)
HRP Diluent	1 (23ml)	1 (23ml)	1 (23ml)	Ready to use
Wash Buffer	1 (10ml)	1 (10ml)	2 (10ml)	200x Concentrate dilute in distilled water (see reagent preparation, section 8)
TMB Substrate	1 (11ml)	1 (11ml)	1 (24ml)	Ready to use
H ₂ SO ₄ stop reagent	1 (11ml)	1 (11ml)	2 (11ml)	Ready to use

4. Materials required but not provided

- Microtitre plate reader fitted with appropriate filters (450nm required with optional 620nm reference filter)
- Microplate washer or wash bottle
- 10, 50, 100, 200 and 1,000 μ l adjustable single channel micropipettes with disposable tips
- 50-300 μ l multi-channel micropipette with disposable tips
- Multichannel micropipette reagent reservoirs
- Distilled water
- Vortex mixer
- Miscellaneous laboratory plastic and/or glass, if possible sterile

5. Storage Instructions

Store kit reagents between 2 and 8°C. Immediately after use remaining reagents should be returned to cold storage (2-8°C). Expiry of the kit and reagents is stated on box front labels. The expiry of the kit components can only be guaranteed if the components are stored properly, and if, in case of repeated use of one component, the reagent is not contaminated by the first handling.

6. Specimen collection, processing & storage

Cell culture supernatants, human serum, plasma or other biological samples will be suitable for use in the assay. Remove serum from the clot or red cells, respectively, as soon as possible after clotting and separation.

Cell culture supernatants: Remove particulates and aggregates by spinning at approximately 1000 x g for 10 min.

Serum: Use pyrogen/endotoxin free collecting tubes. Serum should be removed rapidly and carefully from the red cells after clotting. Following clotting, centrifuge at approximately 1000 x g for 10 min and remove serum.

Plasma: EDTA, citrate and heparin plasma can be assayed. Spin samples at 1000 x g for 30 min to remove particulates. Harvest plasma.

Storage: If not analyzed shortly after collection, samples should be aliquoted (250-500 μ l) to avoid repeated freeze-thaw cycles and stored frozen at -70°C. Avoid multiple freeze-thaw cycles of frozen specimens.

Recommendation: Do not thaw by heating at 37°C or 56°C. Thaw at room temperature and make sure that sample is completely thawed and homogeneous before use. When possible avoid use of badly haemolysed or lipemic sera. If large amounts of particles are present these should be removed prior to use by centrifugation or filtration.

7. Safety & precautions for use

- Handling of reagents, serum or plasma specimens should be in accordance with local safety procedures , e.g.CDC/NIH Health manual : " Biosafety in Microbiological and Biomedical Laboratories" 1984
- The human serum included in this kit have been tested and found non reactive for HbsAg, anti HIV1 & 2 and anti VHC. Nevertheless, no known method can offer complete assurance that human blood derivatives will not transmit hepatitis, AIDS or other infections. Therefore handling of reagents, serum or plasma specimens should be in accordance with local safety procedures
- Laboratory gloves should be worn at all times
- Avoid any skin contact with H_2SO_4 and TMB. In case of contact, wash thoroughly with water
- Do not eat, drink, smoke or apply cosmetics where kit reagents are used
- Do not pipette by mouth
- When not in use, kit components should be stored refrigerated or frozen as indicated on vials or bottles labels
- All reagents should be warmed to room temperature before use. Lyophilized standards should be discarded after use
- Once the desired number of strips has been removed, immediately reseal the bag to protect the remaining strips from deterioration
- Cover or cap all reagents when not in use
- Do not mix or interchange reagents between different lots
- Do not use reagents beyond the expiration date of the kit
- Use a clean disposable plastic pipette tip for each reagent, standard, or specimen addition in order to avoid cross contamination, for the dispensing of H_2SO_4 and substrate solution, avoid pipettes with metal parts
- Use a clean plastic container to prepare the washing solution
- Thoroughly mix the reagents and samples before use by agitation or swirling
- All residual washing liquid must be drained from the wells by efficient aspiration or by decantation followed by tapping the plate forcefully on absorbent paper. Never insert absorbent paper directly into the wells
- The TMB solution is light sensitive. Avoid prolonged exposure to light. Also, avoid contact of the TMB solution with metal to prevent colour development. Warning TMB is toxic avoid direct contact with hands. Dispose off properly
- If a dark blue colour develops within a few minutes after preparation, this indicates that the TMB solution has been contaminated and must be discarded. Read absorbance's within 1 hour after completion of the assay
- When pipetting reagents, maintain a consistent order of addition from well-to-well. This will ensure equal incubation times for all wells
- Follow incubation times described in the assay procedure
- Dispense the TMB solution within 15 min of the washing of the microtitre plate

8. Assay Preparation

Bring all reagents to room temperature before use

8.1. Assay Design

Determine the number of microwell strips required to test the desired number of samples plus appropriate number of wells needed for running zeros and standards. Each sample, standard, zero and control should be tested **in duplicate**. Remove sufficient Microwell Strips for testing from the aluminium pouch immediately prior to use. Return any wells not required for this assay with desiccant to the pouch. Seal tightly and return to 2-8°C storage.

Example plate layout (example shown for a 6 point standard curve)

Standards / Controls		Sample Wells											
		1	2	3	4	5	6	7	8	9	10	11	12
A	2000	2000											
B	1000	1000											
C	500	500											
D	250	250											
E	125	125											
F	62.5	62.5											
G	zero	zero											
H	Ctrl	Ctrl											

All remaining empty wells can be used to test samples in duplicate

8.2. Preparation of Wash Buffer

Dilute the (200x) wash buffer concentrate 200 fold with distilled water to give a 1x working solution. Pour entire contents (10 ml) of the Washing Buffer Concentrate into a clean 2,000 ml graduated cylinder. Bring final volume to 2,000 ml with glass-distilled or deionized water. Mix gently to avoid foaming. Transfer to a clean wash bottle and store at 2°-25°C.

8.3. Preparation of Standard Diluent Buffer

Add the contents of the vial (10x concentrate) to 225ml of distilled water before use.

This Solution can be stored at 2-8°C for up to 1 week.

8.4. Preparation of Standard

Depending on the type of samples you are assaying, the kit may include two standard diluents. Because biological fluids might contain proteases or cytokine-binding proteins that could modify the recognition of the cytokine you want to measure, you should reconstitute standard vials with the most appropriate Standard Diluent.

For **serum and plasma** samples: use Standard Diluent - human serum

For **cell culture supernatants**: use Standard Diluent Buffer

Standard vials must be reconstituted with the volume of standard diluent shown on the vial immediately prior to use. This reconstitution gives a stock solution of 2000pg/ml of IL-8. Mix the reconstituted standard gently by repeated aspirations/ejections. Serial dilutions of the standard are made directly in the assay plate to provide the concentration range from 2000 to 62.5pg/ml. A fresh standard curve should be produced for each new assay.

- Immediately after reconstitution add 200µl of the reconstituted standard to well's A1 and A2, which provides the highest concentration standard at 2000pg/ml
- Add 100µl of appropriate standard diluent to the remaining standard wells B1 and B2 to F1 and F2
- Transfer 100µl from wells A1 and A2 to B1 and B2. Mix the well contents by repeated aspirations and ejections taking care not to scratch the inner surface of the wells
- Continue this 1:1 dilution using 100µl from wells B1 and B2 through to wells F1 and F2 providing a serial diluted standard curve ranging from 2000pg/ml to 62.5pg/ml
- Discard 100µl from the final wells of the standard curve (F1 and F2)

Alternatively these dilutions can be performed in separate clean tubes and immediately transferred directly into the relevant wells.

8.5. Preparation of Controls

Freeze-dried control vials should also be reconstituted with the most appropriate Standard Diluent to your samples.

For **serum and plasma samples**: use Standard Diluent - human serum

For **cells culture supernatants**: use Standard Diluent Buffer

The supplied Controls must be reconstituted with the volume of Standard Diluent indicated on the vial. Reconstitution of the freeze-dried material with the indicated volume, will give a solution at the concentration stated on the vial. Do not store after use.

8.6. Preparation of Biotinylated anti-IL-8

It is recommended this reagent is prepared immediately before use. Dilute the biotinylated anti-IL-8 with the biotinylated antibody diluent in an appropriate clean glass vial using volumes appropriate to the number of required wells. Please see example volumes below:

Number of wells required	Biotinylated Antibody (μ l)	Biotinylated Antibody Diluent (μ l)
16	40	1060
24	60	1590
32	80	2120
48	120	3180
96	240	6360

8.7. Preparation of Streptavidin-HRP

It is recommended to centrifuge vial for a few seconds in a microcentrifuge to collect all the volume at the bottom.

Dilute the 5 μ l vial with 0.5ml of HRP diluent **immediately before use**. Do-not keep this diluted vial for future experiments. Further dilute the HRP solution to volumes appropriate for the number of required wells in a clean glass vial. Please see example volumes below:

Number of wells required	Streptavidin-HRP (μ l)	Streptavidin-HRP Diluent (ml)
16	30	2
24	45	3
32	60	4
48	75	5
96	150	10

9. Method

We strongly recommend that every vial is mixed without foaming prior to use.

Prepare all reagents as shown in section 8.

Note: final preparation of Biotinylated Secondary Antibody (section 8.6) and Streptavidin-HRP (section 8.7) should occur immediately before use.

Assay Step		Details
1.	Addition	Prepare Standard curve as shown in section 8.4 above
2.	Addition	Add 100µl of each, Sample, Standard, Control and zero (appropriate standard diluent) in duplicate to appropriate number of wells
3.	Addition	Add 50µl of diluted biotinylated anti-IL-8 to all wells
4.	Incubation	Cover with a plastic plate cover and incubate at room temperature (18 to 25°C) for 1 hour(s)
5.	Wash	Remove the cover and wash the plate as follows: a) Aspirate the liquid from each well b) Dispense 0.3 ml of 1x washing solution into each well c) Aspirate the contents of each well d) Repeat step b and c another two times
6.	Addition	Add 100µl of Streptavidin-HRP solution into all wells
7.	Incubation	Cover with a plastic plate cover and incubate at room temperature (18 to 25°C) for 30 min
8.	Wash	Repeat wash step 5.
9.	Addition	Add 100µl of ready-to-use TMB Substrate Solution into all wells
10.	Incubation	Incubate in the dark for 12-15 minutes* at room temperature. Avoid direct exposure to light by wrapping the plate in aluminium foil.
11.	Addition	Add 100µl of H₂SO₄:Stop Reagent into all wells
Read the absorbance value of each well (immediately after step 11.) on a spectrophotometer using 450 nm as the primary wavelength and optionally 620 nm as the reference wave length (610 nm to 650 nm is acceptable).		

*Incubation time of the substrate solution is usually determined by the ELISA reader performance. Many ELISA readers only record absorbance up to 2.0 O.D. Therefore the colour development within individual microwells must be observed by the analyst, and the substrate reaction stopped before positive wells are no longer within recordable range

10. Data Analysis


Calculate the average absorbance values for each set of duplicate standards, controls and samples. Ideally duplicates should be within 20% of the mean.

Generate a linear standard curve by plotting the average absorbance of each standard on the vertical axis versus the corresponding IL-8 standard concentration on the horizontal axis.

The amount of IL-8 in each sample is determined by extrapolating OD values against IL-8 standard concentrations using the standard curve.

Example IL-8 Standard curve

Standard	IL-8 Conc	OD (450nm) mean	CV (%)
1	2000	2.154	1.22
2	1000	1.173	1.03
3	500	0.663	2.53
4	250	0.378	5.16
5	125	0.249	4.01
6	62.5	0.174	5.79
zero	0	0.086	7.55



Note: curve shown above should not be used to determine results. Every laboratory must produce a standard curve for each set of microwell strips assayed.

11. Assay limitations

Do not extrapolate the standard curve beyond the maximum standard curve point. The dose-response is non-linear in this region and good accuracy is difficult to obtain. Concentrated samples above the maximum standard concentration must be diluted with Standard diluent or with your own sample buffer to produce an OD value within the range of the standard curve. Following analysis of such samples always multiply results by the appropriate dilution factor to produce actual final concentration.

The influence of various drugs on end results has not been investigated. Bacterial or fungal contamination and laboratory cross-contamination may also cause irregular results.

Improper or insufficient washing at any stage of the procedure will result in either false positive or false negative results. Completely empty wells before dispensing fresh Washing Buffer, fill with Washing Buffer as indicated for each wash cycle and do not allow wells to sit uncovered or dry for extended periods.

Disposable pipette tips, flasks or glassware are preferred, reusable glassware must be washed and thoroughly rinsed of all detergents before use.

As with most biological assays conditions may vary from assay to assay therefore **a fresh standard curve must be prepared and run for every assay.**

12. Performance Characteristics

12.1. Sensitivity

The sensitivity, minimum detectable dose of IL-8 using this Diaclone IL-8 ELISA kit was found to be **29pg/ml**. This was determined by adding 3 standard deviations to the mean OD obtained when the zero standard was assayed 30 times.

12.2. Specificity

The assay recognizes both natural and recombinant human IL-8. To define the specificity of this ELISA several proteins were tested for cross reactivity. There was no cross reactivity observed for any protein tested (IL-1 α , IL-1 β , IL-10 IL-12, IFN γ , IL-2, IL-6, TNF α , IL-4 and IL-13).

12.3. Precision

Intra-assay

Reproducibility within the assay will be evaluated in three independent experiments. Each assay will be carried out with 6 replicates (3 duplicates) of 3 spiked serum or human pooled serum samples and 2 supernatants containing different concentrations of IL-8. **The overall intra-assay coefficient of variation has been calculated to be 3.1% .**

Session	Sample	Assay 1 IL-8 pg/ml	Assay 2 IL-8 pg/ml	Assay 3 IL-8 pg/ml	Mean IL-8 pg/ml	SD	CV %
1	1	1647	1619	1646	1637	16	1.0
	2	985	962	984	977	13	1.3
	3	805	779	719	768	44	5.8
	4	1444	1481	1447	1457	21	1.4
	5	883	864	853	867	15	1.8
2	1	1638	1491	1403	1511	119	7.9
	2	855	873	841	856	16	1.9
	3	684	686	669	680	9	1.4
	4	1300	1262	1312	1291	26	2.0
	5	778	745	788	770	22	2.9
3	1	1599	1468	1496	1521	29	4.5
	2	1019	996	1058	1024	69	3.1
	3	735	668	694	699	31	4.8
	4	1355	1294	1401	1350	34	3.9
	5	792	775	816	794	54	2.6

Inter-assay

Assay to assay reproducibility within one laboratory will be evaluated in three independent experiments by two technicians. Each assay will be carry out with 6 replicates of 3 spiked serum human pooled serum samples and 2 supernatants containing different concentration of IL-8. **The calculated overall coefficient of variation was 9.7%.**

Technician	Session	Sample 1 IL-8 pg/ml	Sample 2 IL-8 pg/ml	Sample 3 IL-8 pg/ml	Sample 4 IL-8 pg/ml	Sample 5 IL-8 pg/ml	
A	1	1874	1166	809	1668	1001	
		1824	1180	852	1690	989	
		1853	1154	875	1632	995	
	2	1929	1090	770	1578	969	
		1784	1089	828	1669	906	
		1796	1111	855	1671	922	
	3	1770	905	762	1594	923	
		1713	898	806	1592	895	
		1734	908	825	1552	946	
B	1	1647	985	805	1444	883	
		1619	962	779	1481	864	
		1646	984	719	1447	853	
	2	1638	855	684	1300	778	
		1491	873	686	1262	745	
		1403	841	669	1312	788	
	3	1599	1019	735	1355	792	
		1468	996	668	1294	775	
		1496	1058	694	1401	816	
Mean IL-8 pg/ml		1682	1004	768	1497	880	
SD		152	111	69	149	83	
CV %		9.0	11.0	8.9	9.9	9.4	

12.4. Dilution Parallelism

Four human pooled serum samples with different levels of IL-8 were analysed at different serial two fold dilutions (1:2 To 1:8) with two replicates each. Recoveries ranged from 77 to 105% with an overall **mean recovery of 88%**.

12.5. Spike Recovery

The spike recovery was evaluated by spiking 3 concentrations of IL-8 in human serum in 3 separate experiments. Recoveries ranged from 96 to 110% with an overall **mean recovery of 102%**

12.6. Stability

Storage Stability

Aliquots of spiked serum samples were stored at -20°C, 2-8°C, room temperature (RT) and at 37°C and the IL-8 level determined after 24h. There was no significant loss of IL-8 reactivity during storage at RT, 2-8°C and 37°C.

Freeze-thaw Stability

Aliquots of spiked serum were stored frozen at -20°C and thawed up to 5 times and the IL-8 level was determined. There was a decrease in activity of IL-8 after 5 cycles of freezing and thawing.

12.7. Expected serum values

A panel of 20 human sera was tested for IL-8. 18 were below the detection level of 29pg/ml. Two samples reported results of 143pg/ml and 197pg/ml.

12.8. Standard Calibration

This immunoassay is calibrated against the International Reference Standard NIBSC 89/520. NIBSC 89/520 is quantitated in International Units, and equivalence in ng/ml is indicated. 1ng NIBSC corresponding to 1ng Diaclone IL-8.

13. Bibliography

- Wolff b, Burns AR, Middleton J, Rot A. J. Exp. Med. 1998 Nov 2;188(9):1757-62
Utggaard JO, Jahnsen FL, Bakka A, Brandtzaeg P, Haraldsen G. J. Exp. Med. 1998 Nov 2;188(9):1751-6
Bagnolini M, Clark-Lewis I (1992). FEBS Lett. 307 (1):97-101
Matsushima, K. et al (1988) J. Exp. Med. 167, 1883-1893.
Peichl P., M. Ceska, H.Broell, F. Effenberger, and I. J. D. Lindley. (1992) Annals of the Rheumatic Diseases 51, 19-22.
Elford, P. R. , and P. H. Cooper . (1991) Arthritis and Rheumatism 34, 325-332
Gearing, A. J. H. , N. J. Fincham, C. R. Bird, M. Wadhwa, A. Meager, J. E. Cartwright, and R. D. R. Camp. (1990) Cytokine 2, 68-75
Matsushima, K. and J.J. Oppenheim. (1989). Cytokine 1, 2-13.

14. Diaclone Human IL-8 ELISA references

- Audibert, C. et al., Infect Immun., 2001; 69(3): 1625-9.
Backhed F. et al., J. Biol. Chem., 2002; 277(20): 18198 - 19205
Backhed, F. et al., Infect Immun., 2003; 71(6): 3357-60.
Bonnans, C. et al., Am J Pathol., 2006 ; 168 (4) : 1064-72.
Chabbert-de Ponnat, I. et al., Int Immunol., 2005; 17(4): 439-47.
Coconnier, M. H. et al., Appl Environ Microbiol., 2000;66(3): 1152-7.
Ellmerich S. et al., Carcinogenesis, 2000; 21(4): 753 -756
Gironella J. et al., Gut, 2005; 54(9): 1244 - 1253
Krzysiek, R. et al., J Immunol., 1999;162(8): 4455-63
Lekkou A. et al., Clin. Diagn.Lab. Immunol., 2004; 11(1):161 - 167
Rajappa, M. et al., Angiology, 2009; 60(4): 419-426
Saraiva M. et al., JEM 2002; 196(6): 829-839
Sell, H. et al., Am J Physiol Endocrinol Metab., 2008;294:E1070-E1077
Sjolinder,H et al., Infect Immun., 2008; 76(9)3959-66
Tenca C. et al., 2005; 174(11): 6757 - 6763
van de Sande W.W.J. et al., J. Immunol., 2007; 179(5): 3065-3074.
Yadav, U. C. S. et al., J. Immunol., 2009; 183(7): 4723-4732
Zrioual, S. et al., J. Immunol., 2008; 180(1): 655-663.

15. Assay Summary

Total procedure length: 1h45mn

Add 100 µl of sample and diluted standard/controls and 50µl Biotinylated anti-IL-8



Incubate 1 hours at room temperature



Wash three times



Add 100µl of Streptavidin-HRP



Incubate 30min at room temperature



Wash three times



Add 100µl of ready-to-use TMB

Protect from light. Let the color develop for 12-15 mn.



Add 100µl H₂SO₄



Read Absorbance at 450 nm

Products Manufactured and Distributed by:

Diaclone SAS
1 Boulevard A.Fleming
25020 Besancon Cedex
France
Tel +33 381413838
Fax +33 381413636
Email: info@diacalone.com

16. International Summaries

16.1. French

PREPARATION DES REACTIFS : RESUME

1. Tampon de Lavage (Washing Buffer)

Ajouter 10 ml de **Tampon de Lavage concentré** (*Washing Buffer Concentrate*) 200 fois (200X) à 1990 ml d'eau distillée

2. Tampon de Dilution du Standard (Standard Diluent Buffer)

Ajouter 25 ml de **Tampon de Dilution du Standard concentré** 10 fois (*Standard Diluent Buffer Concentrate 10X*) à 225 ml d'eau distillée

3. Standard IL-8 (IL-8 Standard)

Reconstituer le **Standard IL-8** (*IL-8 Standard*) en ajoutant la quantité indiquée sur le flacon avec le **Tampon approprié** (Tampon de Dilution du Standard ou Tampon de Dilution du Standard : Sérum Humain)

4. Contrôles (Controls)

Reconstituer les **Contrôles** (*Controls*) en ajoutant la quantité indiquée sur le flacon avec le **Tampon approprié** (Tampon de Dilution du Standard ou Tampon de Dilution du Standard : Sérum Humain)

5. Anti-IL-8 Biotinylé (Biotinylated anti IL-8)

Nombre de barrettes	Anti-IL-8 Biotinylé Concentré (µl)	Diluent de l'Anticorps Biotinylé (µl)
2	40	1060
3	60	1590
4	80	2120
6	120	3180
12	240	6360

6. Streptavidin-HRP (Streptavidin-HRP)

Nombre de barrettes	Streptavidin-HRP pré-diluée (µl)	Diluent HRP (ml)
2	30	2
3	45	3
4	60	4
6	75	5
12	150	10

RESUME DU PROTOCOLE OPERATOIRE: durée totale : 1h45mn

1. Ajouter 100 µl de **Tampon de Dilution du Standard** (*Standard Diluent Buffer*) approprié, en duplicate, dans les puits Standards (B1 à F2)
2. Ajouter à la pipette 200 µl de **Standard IL-8** (*IL-8 Standard*) reconstitué dans les puits A1 et A2 puis réaliser des dilutions du Standard allant de 2000 à 62.50 pg/ml en transférant 100 µl d'un puit à l'autre. Jeter les 100 µl des derniers puits (F1 et F2) .
3. Ajouter 100 µl de **Tampon de Dilution du Standard** approprié (*Standard Diluent Buffer ou Standard Diluent : human serum*) en duplicate dans les puits "blancs".
4. Ajouter 100 µl d'**échantillon** (*Sample*), en duplicate, dans les puits désignés et 100 µl de **contrôle** (*control*), en duplicate dans les puits contrôles.
5. Préparer l'**anticorps anti IL-8 Biotinylé** (*Biotinylated anti IL-8*)
6. Ajouter 50 µl d'**anticorps anti IL-8 Biotinylé dilué** (*diluted biotinylated anti IL-8*) dans tous les puits.
7. Couvrir les barrettes de puits et incuber pendant 1 heure à température ambiante (18-25°C)
8. Vider et laver les puits 3 fois avec le **Tampon de Lavage** (*Washing Buffer*)
9. Préparer la Streptavidin-HRP
10. Ajouter 100 µl de **Streptavidin-HRP diluée** (*diluted HRP-Streptavidin*) dans tous les puits
11. Couvrir les puits et incuber pendant 30 minutes à température ambiante (18°-25°C).
12. Vider et laver les puits 3 fois avec le **Tampon de Lavage** (*Washing Buffer*)
13. Ajouter 100 µl de solution de TMB (*TMB solution*) prête à l'emploi dans tous les puits y compris les "blancs".
14. Incuber les barrettes de puits pendant environ 12-15 minutes à température ambiante (18°-25°C) à l'obscurité.
15. Ajouter 100 µl d'**H₂SO₄: Solution Stop** (*H₂SO₄: Stop Solution*) dans tous les puits y compris les "blancs".
16. Mesurer l'absorbance à la longueur d'onde 450 nm et optionnellement à 620 nm (entre 610 et 650 nm) comme longueur d'onde de référence.

Remarque: Les échantillons présentant une valeur de D.O. excédant la gamme de la courbe Standard peuvent résulter à des taux d'IL-8 incorrects. C'est pourquoi, il est recommandé de diluer de tels échantillons avec le **Tampon de Dilution du Standard** (*Standard Diluent Buffer*) afin de quantifier précisément le véritable taux d'IL-8.

16.2. Spanish

PREPARACIÓN DE LOS PRODUCTOS

1. Tampón de Lavado	Añadir Tampón de Lavado Concentrado 200 X (10 ml) a 1990 ml de agua destilada.		
2 Tampón diluyente del estándar	Añadir Tampón Diluyente del Estándar Concentrado 10 X (25 ml) a 225 ml de agua destilada.		
3. Estándar IL-8	Reconstituir el Estándar IL-8 añadiendo el Diluyente del Estándar apropiado, como indica la etiqueta del vial.		
4. Controles	Reconstituir los controles añadiendo el Diluyente del Estándar apropiado, como indica la etiqueta del vial.		
5. Anti-IL-8 biotinilado	Número de tiras	Anticuerpo biotinilado concentrado (μ l)	Diluyente del anticuerpo biotinilado (μ l)
	2	40	1,060
	3	60	1,590
	4	80	2,120
	6	120	3,180
	12	240	6,360
6. Estreptavidina-HRP	Número de tiras	Estreptavidina-HRP prediluida (μ l)	Diluyente de HRP (ml)
	2	30	2
	3	45	3
	4	60	4
	6	75	5
	12	150	10

RESUMEN DEL PROTOCOLO. El procedimiento total tiene una duración de 1h45min.

1. Añadir 100 μ l del **Tampón Diluyente del Estándar** apropiado, por duplicado, a los pocillos designados para el estándar (B1 to F2).
2. Pipetear 200 μ l del **Estándar IL-8** reconstituido en los pocillos A1 y A2 y hacer diluciones seriadas del estándar con el rango de concentraciones de 2000 al 62.50 pg/ml, transfiriendo 100 μ l de un pocillo al siguiente. Descartar 100 μ l de los últimos pocillos.
3. Añadir 100 μ l del **Tampón Diluyente del Estándar** apropiado, por duplicado, a los pocillos que van a ser el “blanco”.
4. Añadir 100 μ l de las muestras, por duplicado, a los pocillos designados para ello, y 100 μ l del Control reconstituido, por duplicado, a los pocillos designados como “control”.
5. Preparar el anticuerpo **Anti-IL-8 Biotinilado**.
6. Añadir 50 μ l del **anti-IL-8 Biotinilado** y diluido, a todos los pocillos.
7. Cubrir la placa e incubar durante 1 horas a temperatura ambiente (18-25°C).
8. Vaciar y lavar la placa 3 veces con **Tampón de Lavado**.
9. Preparar la **Estreptavidina-HRP**.
10. Añadir 100 μ l de **Estreptavidina-HRP** diluida a todos los pocillos.
11. Cubrir la placa e incubar 30 minutos a temperatura ambiente (18° to 25°C).
12. Vaciar y lavar la placa 3 veces con **Tampón de Lavado**.
13. Añadir 100 μ l de solución **TMB preparado para utilizar**, a todos los pocillos, incluidos los pocillos con “blancos”.
14. Incubar la placa durante 12-15 minutos a temperatura ambiente (18° to 25°C) y en oscuridad.
15. Añadir 100 μ l de H₂SO₄: **Solución de Parada**, a todos los pocillos, incluidos los pocillos con los “blancos”.
16. Medir la intensidad de color a 450 nm y a 620 nm como longitud de onda de referencia (de 610 nm a 650 nm sería aceptable).

Nota: El cálculo de concentraciones de muestras con densidad óptica que supere el rango de la curva estándar, resultaría incorrecto, dando niveles de IL-8 más bajos de lo real. Estas muestras, requerirían ser diluidas con el Tampón de Dilución de Estándar adecuado, para poder precisar la cantidad real de IL-8 .